

ASTM D 2887 Extended Standard Test Method for Boiling Range Distribution of Petroleum Fractions by Gas Chromatograph

DCG has available a gravimetrically prepared NIST Traceable by weight calibration standard for the D 2887 Extended method. This method covers the boiling range of normal alkanes from C₅ – C₆₀. DCG's calibration standard for D 2887 Extended in addition to being gravimetrically prepared and NIST Traceable by weight has been verified by one or more analytical methods for component retention time. The addition of 2 mls of CS₂ is required to dilute this standard before analysis. Custom standards are available upon request.

Part #: D 2887 Extended : \$50.00

Concentration: Weight

Packaged in 2 mL GC vials.

4.35% N-Pentane
4.35% N-Hexane
4.35% N-Heptane
4.35% N-Octane
4.35% N-Nonane
4.35% N-Decane
4.35% N-Undecane
4.35% N-Dodecane
4.35% N-Tetradecane
4.35% N-Pentadecane
4.35% N-Hexadecane
4.35% N-Heptadecane
4.35% N-Octadecane
4.35% N-Eicosane
4.35% N-Tetracosane
4.35% N-Octacosane
4.35% N-Dotriacontane
4.35% N-Hexatriacontane
4.35% N-Tetracontane
4.35% N-Tetratetracontane
4.35% N-Pentacontane
4.35% N-Tetrapentacontane
4.35% N-Hexacontane

Prices are subject to change

4170 A Main Pearland, TX 77581 PH: 281-648-1894 FAX: 281-648-1895
www.dcgpartnership.com
E-mail: sales@dcgpartnership.com

Chromatographic Reference Materials