

ASTM D 6730 System and Column Evaluation Mixture for the Determination of Individual Components in Spark Ignition Engine Fuels by 100-Metre Capillary (with Pre-column) High-Resolution Gas Chromatography

DCG has available an NIST Traceable by weight gravimetrically prepared calibration standard for ASTM D 6730 system and column evaluation. This standard in addition to being gravimetrically prepared and NIST Traceable by weight has been verified by one or more analytical methods. Custom standards are available upon request.

Part #: ASTM D 6730 SCE: \$100.00

Packaged in 2 ml pre-scored ampoule

8.00% Ethanol
2.00% N-Pentane
0.50% t-Butanol
2.50% 2-Methyl-2-Butene
0.50% 2, 3 Dimethylbutane
10.0% Methyl Tertiary Butyl Ether
2.00% N-Hexane
0.50% 1-Methylcyclopentene
1.00% Benzene
28.9% Cyclohexane
0.20% 3-Ethylpentane
0.50% Trans -1, 2-Dimethylcyclopentane
2.00% N-Heptane
0.50% 2, 3, 3-Trimethylpentane
7.00% Toluene
2.00% N-Octane
25.0% Ethylbenzene
1.00% p-Xylene
0.20% 2, 3-Dimethylheptane
2.00% N-Nonane
0.20% 5-methylnonane
0.50% 1-Methyl-2-Ethylbenzene
1.00% N-Decane
0.50% Undecane
0.25% 1, 2, 3, 5-Tetramethylbenzene
0.50% Naphthalene
0.25% Dodecane
0.25% 1-Methylnaphthalene
0.25% Tridecane

Prices are subject to change

4170 A Main Pearland, TX 77581 PH: 281-648-1894 FAX: 281-648-1895

www.dcgpartnership.com

E-mail: sales@dcgpartnership.com

Chromatographic Reference Materials