



## **Standard Test Method for Cloud Point of Petroleum Products in Accordance with ASTM D 2500**

DCG has available Cloud Point Reference Standards for Petroleum Products and Biofuels that are transparent in layers 40 mm. in thickness and with a cloud point below 49° C for compliance with ASTM D 2500. The cloud point values have been determined by consensus analysis. The following cloud point standards are available in 200 mL bottles.

**Part #: DCGCP 5**

Description: Approximate Cloud Point value +5° C\*

**Part #: DCGCP 2**

Description: Approximate Cloud Point value -2° C\*

**Part #: DCGCP10**

Description: Approximate Cloud Point value -10° C\*

**Part #: DCGCP15**

Description: Approximate Cloud Point value -15° C\*

**Part #: DCGCP20**

Description: Approximate Cloud Point value -20° C\*

\*Certified values provided with standards