



Standards for the Determination of Organic Chlorides in Petroleum Liquids to be used with ASTM D 4929 And ASTM D 5808

DCG has available gravimetrically prepared N.I.S.T. traceable by weight calibration standards for the determination of organic chlorides by Microcoulometry. All standards in addition to being gravimetrically prepared and NIST traceable by weight have been verified by one or more analytical methods. Custom standards are available.

Part #: CLISO Cal Kit #1 5.00 ppm – 100 ppm: \$120.00

Calibration Kit for Total Chloride by weight from 1-Chloropentane in Isooctane for the 5.00 -100 ppm range. Packaged in 2 ml. ampoules.

Blank

- 5.00 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane
- 10.0 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane
- 25.0 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane
- 50.0 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane
- 100 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane

Part #: CLISO Cal Kit #2 1.00 ppm- 25.0 ppm: \$120.00

Calibration Kit for Total Chloride by weight from 1-Chloropentane in Isooctane for the 1.00 – 25.0 ppm range. Packaged in 2 ml. ampoules.

Blank

- 1.00 w/w ppm Total Chlorine from 1-Chloropentane Isooctane
- 5.00 w/w ppm Total Chlorine from 1-Chloropentane Isooctane
- 10.0 w/w ppm Total Chlorine from 1-Chloropentane Isooctane
- 15.0 w/w ppm Total Chlorine from 1-Chloropentane Isooctane
- 25.0 w/w ppm Total Chlorine from 1-Chloropentane Isooctane

CLISO Chk Std # 1: \$100.00

5 x 2 ml. ampoules - 75.0 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane

CLISO Chk Std # 2: \$100.00

5 x 2 ml. ampoules – 40.0 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane

CLISO Chk Std # 3: \$100.00

5 x 2 ml. ampoules - 10.0 w/w ppm Total Chlorine from 1-Chloropentane in Isooctane

Prices are subject to change